

AGGREGATE

SAND EQUIVALENT SHAKER (AG-040/SS)

ASTM D2419 • EN 933-8

- Provides a constant, uniform and precise shaking for the sand sample in the cylindrical measure used in Sand Equivalent Test Set.
- The equipment has a stroke of $200 \text{ mm} \pm 10 \text{ mm}$, shaking at 175 cycle per minute.
- Equipped with a digital stroke indicator.

**TECHNICAL SPECIFICATIONS**

- Stroke : $200 \text{ mm} \pm 10 \text{ mm}$
- Rate : 175 cpm
- Digital Stroke Indicator
- 220 - 240 V / 50 - 60 Hz