

■ SIEVE SEGREGATION TEST SET (B-023) EN 12350-11

- Used to determine the sieve segregation resistance of Self-Consolidating Concrete (SCC).



● COMPOSED OF

- Perforated Sieve - 5 mm Aperture (Ø 300 mm)
- Pan (Ø 300 mm)
- Bucket (> 11 liters)

■ CONCRETE BLEEDING TEST SET (B-028) EN 480-4

- Used to determine the relative quantity of mixing water that will bleed from a sample of freshly mixed concrete



● COMPOSED OF

- Cylindrical container with Cover (Ø 250 x 280 mm)
- Tamping Rod B-020/G/TR
- Pipette (50 ml) GL-27/50
- Glass Beaker (100 ml) GCB/0100