

CEMENT

WORKABLE LIFE and CORRECTION TIME APPARATUS (C-096)
EN 1015-9

- Used to determine the workable life and correction time of fresh mortar.



TECHNICAL SPECIFICATIONS

- Penetration rod (Ø 5 mm / Ø 6.175 mm x 65 mm)
- Washer (Ø 20 mm)
- Sample container (Ø 75 mm)

PLUNGER PENETRATION APPARATUS (C-094)
EN 413-2 • EN 459-2 • EN 1015-4

- Used to determine the consistency of cement mix.



COMPOSED OF

- Steel base and frame
- Test cup (Ø 80 mm x 70 mm)
- Plunger assembly (90 g, dropping from 100 mm)
- Tamper (C-021/T/EN) (Ø 40 mm x 200 mm , 0.250 kg)