

FLASH and FIRE POINT APPARATUS (Pensky-Martens Closed Cup)

Code : A473



- Used to determine the flash point of petroleum products by the Closed Cup Test, with a Flash Point between 40°C to 360°C.

STANDARDS

EN 22719 • ASTM D93 • AASHTO T73

ORDERING INFORMATION

Item	Code
FLASH and FIRE POINT APPARATUS (Pensky-Martens Closed Cup)	A473H00XU