

CEMENT > WORKABILITY & CONSISTENCY

PLUNGER PENETRATION APPARATUS

Code : C094



- Used to determine the consistency of cement mix.

STANDARDS

EN 413-2 • EN 459-2 • EN 1015-4

SET

- Steel base and frame
- Test cup (Ø 80 mm x 70 mm)
- Plunger assembly (90 g, dropping from 100 mm)
- Tamper (Ø 40 mm x 200 mm , 0.250 kg)

ORDERING INFORMATION

Item	Code
PLUNGER PENETRATION APPARATUS	C094X00EH