

SOIL > CONSOLIDATION

OEDOMETER - CONSOLIDATION

Code: T210



- Used to determine the behavior of the soil sample upon certain loading in a specified period of time. It indicates the settlement characteristics of the soil which is known as Consolidation.
- The loading ratio for the device is 10:1 and designed as front loading type. The rear weight is used for balancing the device.
- The device has a support for a screw jack and supplied with a dial and its holder, consolidation cell, two porous discs and weight set.
- Fixed on a frame and can be ordered as 1, 2 or 3 devices on the same frame.

STANDARDS

ASTM D2435 • ASTM D3877 • ASTM D4546 • BS 1377-5



TECHNICAL SPECIFICATIONS

- Front Loading Type
- Load Ratio = 9:1 10:1 11:1
- Rear Balancing Weight
- Screw Jack Support

SUPPLIED WITH

- Dial Indicator: 12.7 / 0.002 mm
- Dial Holder
- Consolidation Cell
- Porous Discs (2 ea)
- Weight Set (3 x 10 kg, 2 x 5 kg, 3 x 2 kg, 3 x 1 kg, 2 x 0.5 kg)

ORDERING INFORMATION

| ltem | Code |
|--|-----------|
| AUTOMATIC CONSOLIDATION DEVICE - 10 kN | T210A010U |
| OEDOMETER - CONSOLIDATION | T210H00XH |
| CONOLIDATION CELL - Ø 50 mm | T210P001H |
| CONOLIDATION CELL - Ø 63.50 mm | T210P002H |
| CONOLIDATION CELL - Ø 75 mm | T210P003H |
| FRAME for OEDOMETER (1 Unit) | T210P004H |
| FRAME for OEDOMETER (2 Units) | T210P005H |





| FRAME for OEDOMETER (3 Units) | T210P006H |
|--|-----------|
| SUCTION-CONTROLLED OEDOMETER CELL | T210P007H |
| SUCTION-CONTROLLED OEDOMETER CELL with DIAMETRIC CALIPER | T210P008H |
| PRE-COMPRESSION FRAME for 3 SPECIMENS | T210P009H |
| OEDOMETER - CONSOLIDATION (With Electronic Def. Sensor) | T211X00XH |
| WEIGHT SET - 50 kg | T215X050H |
| WEIGHT SET - 64 kg | T215X064H |
| WEIGHT SET - 80 kg | T215X080H |
| | |

OTHER PHOTOS



